



Chapter XIII

3G Mobile Medical Image Viewing

Eric T. T. Wong, The Hong Kong Polytechnic University, Hong Kong
Carrison K. S. Tong, Pamela Youde Nethersole Eastern Hospital, Hong Kong

Abstract

Tele-radiology is the technology of remote medical consultation using X-ray, computed tomographic, or magnetic resonance images. It was commonly accepted by clinicians for its effectiveness in making diagnoses for patients in critical situations. Because of the huge size of data volume involved in tele-radiology (American College of Radiology [ACR], 2003), clinicians are not satisfied with the relatively slow data-transfer rate. It limits the technology to fixed-line communication between the doctor's home and his or her office. In this project, a mobile high-speed wireless medical image viewing system using a 3G (third-generation) wireless network (Collins & Smith, 2001), virtual private network, and one-time two-factor authentication (OTTFa) technologies is presented. Using this system, tele-radiology can be achieved by using a 3G PDA (personal digital assistant) phone to query, retrieve, and review the patient's record at anytime and anywhere in a secure environment. Using this technology, the patient-data availability can be improved significantly, which is crucial to timely diagnoses of patients in critical situations.

Introduction

Tele-radiology is the technology of remote medical consultation using X-ray, computed tomographic (CT), or magnetic resonance (MR) images. This technique is commonly accepted by clinicians for its effectiveness in making diagnoses for patients in critical situations. For effective implementation of tele-radiology, many technical problems including data integrity, accessibility, size of data volume, compression method, and bandwidth of linkage should be considered. Hitherto, due to the huge size of data volume involved, clinicians are not satisfied with the slow data-transfer rate. It limits the use of the technology to fixed-line communication between a doctor's office and his or her home. In this project, a mobile high-speed wireless medical image viewing system using third-generation (3G) mobile technology, a virtual private network (VPN), common gateway interfacing (CGI), dynamic JPEG (Joint Photographic Experts Group) compression, the Web, the structural query language (SQL; DuBois, 2002), digital imaging and communication in medicine (DICOM; National Electrical Manufacturers Association [NEMA], 2004), and one-time two-factor authentication (OTTFA) technologies were developed. Using this system, tele-radiology has been enhanced to a large extent; image-data query and retrieval can be transferred from a hospital data centre to any notebook personal computer (PC), or to any 3G personal digital assistant (PDA) phone at anytime and anywhere in a secure environment. Hence, the patient-data availability can be improved significantly, which is quite important for patients in critical situations.

Background

Tele-radiology involves the process of sending radiographic images from one point to another through digital standard telephone lines, wide area networks (WANs), or over a local area network (LAN). The radiographic images can be acquired either by a video-capture board such as a frame grabber or the console of a medical imaging modality. After acquisition, the images were digitally stored in a tele-radiology workstation in which the images were ready to be sent to a remote site over a network such as an Ethernet.

In the field of medical imaging, most of the images were stored in DICOM formation. DICOM is a standard that is a framework for medical imaging communication. It was developed by the American College of Radiology (ACR) and the National Electrical Manufacturers Association (Bidgood & Horii, 1992) with input from various vendors, academia, and industry groups. It is referred to as Version 3.0 because it replaces Versions 1.0 and 2.0 of the standard previously issued by ACR

14 more pages are available in the full version of this document, which may be purchased using the "Add to Cart" button on the publisher's webpage: www.igi-global.com/chapter/mobile-medical-image-viewing/31162

Related Content

Electrocardiographic Signal Processing Applications in Telemedicine

Adrian Brezulianu, Iulian Ciocoiuand Monica Fira (2010). *Handbook of Research on Developments in E-Health and Telemedicine: Technological and Social Perspectives* (pp. 219-246).

www.irma-international.org/chapter/electrocardiographic-signal-processing-applications-telemedicine/40649

Phase Unwrapping Using Energy Minimization Methods for MRI Phase Image

Kusworo Adi, Tati L. R. Mengko, Andriyan B. Suksmonoand H. Gunawan (2010). *International Journal of E-Health and Medical Communications* (pp. 50-56).

www.irma-international.org/article/phase-unwrapping-using-energy-minimization/46060

ANN based Expert System to Predict Disease in Cardiac Patients at Initial Stages

Umer Rashid (2015). *International Journal of E-Health and Medical Communications* (pp. 1-9).

www.irma-international.org/article/ann-based-expert-system-to-predict-disease-in-cardiac-patients-at-initial-stages/128297

Use-Case Driven Approach for a Pragmatic Implementation of Interoperability in eHealth

Karima Bourquardand Alexander Berler (2017). *International Journal of Reliable and Quality E-Healthcare* (pp. 52-62).

www.irma-international.org/article/use-case-driven-approach-for-a-pragmatic-implementation-of-interoperability-in-ehealth/181630

The Patient is Dead: Continuing Medical Education and the Hidden Curriculum

Shaifali Bansal (2012). *International Journal of User-Driven Healthcare* (pp. 70-75).

www.irma-international.org/article/patient-dead-continuing-medical-education/64332