## Chapter 80 Brain-Machine Interface Using Brain Surface Electrodes: Real-Time Robotic Control and a Fully Implantable Wireless System

Masayuki Hirata Osaka University Medical School, Japan

**Takufumi Yanagisawa** Osaka University Medical School, Japan

Kojiro Matsushita Osaka University Medical School, Japan

Hisato Sugata Osaka University Medical School, Japan

Yukiyasu Kamitani ATR Computational Neuroscience Laboratories, Japan

> **Takafumi Suzuki** National Institute of Information and Communications Technology, Japan

Hiroshi Yokoi The University of Tokyo, Japan

Tetsu Goto Osaka University Medical School, Japan

Morris Shayne Osaka University Medical School, Japan

Youichi Saitoh Osaka University Medical School, Japan

Haruhiko Kishima Osaka University Medical School, Japan

Mitsuo Kawato ATR Computational Neuroscience Laboratories, Japan

**Toshiki Yoshimine** Osaka University Medical School, Japan

## ABSTRACT

The brain-machine interface (BMI) enables us to control machines and to communicate with others, not with the use of input devices, but through the direct use of brain signals. This chapter describes the integrative approach the authors used to develop a BMI system with brain surface electrodes for real-time robotic arm control in severely disabled people, such as amyotrophic lateral sclerosis patients. This integrative BMI approach includes effective brain signal recording, accurate neural decoding, robust robotic control, a wireless and fully implantable device, and a noninvasive evaluation of surgical indications.

DOI: 10.4018/978-1-4666-4422-9.ch080

## INTRODUCTION

The brain-machine interface (BMI) is a manmachine interface that enables us to control machines and to communicate with others not with the use of input devices, but through the direct use of brain signals alone (Figure 1). Several diseases and conditions can lead to the loss of muscular control without a disruption in patients' brain function, including amyotrophic lateral sclerosis (ALS), brainstem stroke, spinal cord injury, and muscular dystrophy, among others. BMI technology offers these patients greater independence and a higher quality of life by enabling the control of external devices to communicate with others and the ability to manipulate their environment at will (Wolpaw, Birbaumer, McFarland, Pfurtscheller, & Vaughan, 2002).

There are two types of BMI: invasive BMI and noninvasive BMI. Invasive BMI requires surgical procedures and measures the brain signals from intracranial electrodes (needle electrodes or brain surface electrodes), whereas noninvasive BMI measures brain signals noninvasively from outside of the body using scalp electrodes, and so forth. To achieve a higher performance and a higher level of usefulness, we employed invasive BMI techniques, which involve the implantation of devices. For use in a practical situation, invasive BMI requires an organic integration of the following medical and engineering technologies:

- 1. Neural recording with high spatiotemporal resolution.
- 2. High-speed data transfer and processing.
- 3. Optimal extraction of appropriate neurophysiological features.
- 4. Accurate neural decoding.
- 5. Robust control of external devices such as robotic arms and electric wheelchairs.
- 6. Downsizing, integration, and implantation of electronic devices, and the use of wireless technology.

- 7. Noninvasive pre-surgical evaluations for appropriate surgical indications.
- 8. On-target survey and analysis of patient needs.
- 9. Addressing of neuroethical issues.

In this chapter, we describe the development of our invasive BMI system using brain surface electrodes.

## NEURAL DECODING AND REAL-TIME ROBOTIC CONTROL USING ELECTROCORTICOGRAMS

## Clinical Studies Using Electrocorticograms Recorded from Brain Surface Electrodes

In the process of providing neurosurgical treatments for specific groups of patients, we sometimes record brain signals (electrocorticograms: ECoGs) or electrically stimulate the brain using

*Figure 1. A conceptual diagram of the brain machine interface* 



12 more pages are available in the full version of this document, which may be purchased using the "Add to Cart" button on the publisher's webpage: www.igi-global.com/chapter/brain-machine-interface-using-brain-surfaceelectrodes/80687

## **Related Content**

#### Nascent Access Technologies for Individuals with Severe Motor Impairments

Sarah Power, Saba Moghimi, Brian Nhanand Tom Chau (2014). *Assistive Technologies: Concepts, Methodologies, Tools, and Applications (pp. 720-739).* www.irma-international.org/chapter/nascent-access-technologies-for-individuals-with-severe-motor-impairments/80640

#### Improving Students' Academic Learning by Helping Them Access Text

Michael Ben-Avie, Régine Randall, Diane Weaver Dunneand Chris Kelly (2015). *Recent Advances in Assistive Technologies to Support Children with Developmental Disorders (pp. 217-236).* www.irma-international.org/chapter/improving-students-academic-learning-by-helping-them-access-text/131336

#### Role of Assistive Technology in Teaching Students With Disabilities in K-12 Classrooms

Pankaj Khazanchiand Rashmi Khazanchi (2022). Technology-Supported Interventions for Students With Special Needs in the 21st Century (pp. 149-176).

www.irma-international.org/chapter/role-of-assistive-technology-in-teaching-students-with-disabilities-in-k-12classrooms/300026

# A Facial Expression Mediated Natural User Interface Communication Model for Children with Motor Impairments

Ong Chin Annand Lau Bee Theng (2014). Assistive Technologies and Computer Access for Motor Disabilities (pp. 254-284).

www.irma-international.org/chapter/facial-expression-mediated-natural-user/78430

#### Ventricular Assist Device and Its Necessity for Elderly Population

Md. Shamsul Arefin, Nasser K. Awad, Chandra Prakash Rathore, Anupam Shuklaand Yosry S. Morsi (2016). *Optimizing Assistive Technologies for Aging Populations (pp. 314-334).* www.irma-international.org/chapter/ventricular-assist-device-and-its-necessity-for-elderly-population/137799