Chapter 23

Quality of Service and Radio Management in Biomedical Wireless Sensor Networks

Carlos Abreu

Instituto Politécnico de Viana do Castelo, Portugal

Francisco Miranda

Instituto Politécnico de Viana do Castelo and CIDMA of University of Aveiro, Portugal

Paulo M. Mendes

Universidade do Minho, Portugal

ABSTRACT

Biomedical wireless sensor networks enable the development of real time patient monitoring systems, either to monitor chronically ill persons in their homes or to monitor patients in step-down hospital units. However, due to the critical nature of medical data, these networks have to meet demanding quality of service requirements, ensuring high levels of confidence to their users. These goals depend on several factors, such as the characteristics of the network deployment area or the network topology. In such context, this chapter surveys the main applications of biomedical wireless sensor networks, taking into account the key quality of service requirements of each one of them. Finally, it presents an analytic method, and its experimental validation, to help engineers managing the radio power of the network nodes in order to improve the communications and the quality of service provided by the network while minimising the energy consumption and, thus, maximising the network lifetime.

INTRODUCTION

Healthcare providers and professionals broadly use information and communication technologies in their daily practice. Indeed, it is practically impossible to find a healthcare service that does not use any kind of computer-based technology. The information and communication technologies become almost completely pervasive and ubiquitous. The last frontier envisioned by the research community concerns

DOI: 10.4018/978-1-4666-8823-0.ch023

closed-loop real-time and continuous monitoring of each person's health in every aspect of its daily life, using non-intrusive and ubiquitous technologies. Although this futuristic vision is still far from reality, some big steps are being taken in that direction. It is expected that pervasive and ubiquitous healthcare (u-health) systems, combined with wireless sensing systems like Biomedical Wireless Sensor Networks (BWSNs) and Body Sensor Networks (BSNs), contribute to change the actual healthcare practice centred in the episodic evaluation of the patients, to the continuous patient assessment, based on real-time and long-term monitoring.

BWSNs are small-sized Wireless Sensor Networks (WSNs) equipped with biomedical sensors, designed for medical applications or healthcare services. Typical applications of BWSNs include catastrophe and emergency response, Ambient Assisted Living (AAL) applications to monitor and assist disabled or elderly people, and patient monitoring systems for chronically ill persons. Among these application fields, this chapter will focus on the aspects related to the Quality of Service (QoS) guarantees requested to BWSNs by patient monitoring systems used to collect vital and physiological signs of patients in step-down hospital units of nursing homes.

Typical patient monitoring applications, supported by BWSNs, are used to collect vital or physiological signs, such as respiratory rate, pulse rate, temperature, blood pressure, and blood oxygen saturation in order to complement the measurements performed manually by nursing professionals a few times a day and thus, enhancing the quality of the health care provided to patients. The sensed data are then sent to a local or remote database to be used to support healthcare professionals in their medical practice. As an abstraction, BWSNs can be seen as a physical layer for Healthcare Information System (HIS), collecting data to support healthcare professionals in their decisions and medical diagnosis, see Figure 1.




Figure 1. Typical applications of biomedical wireless sensor networks and their architecture

20 more pages are available in the full version of this document, which may be purchased using the "Add to Cart" button on the publisher's webpage:

www.igi-global.com/chapter/quality-of-service-and-radio-management-in-biomedical-wireless-sensor-networks/137460

Related Content

Large-Scale Social Simulation, Dealing with Complexity Challenges in High Performance Environments

C. Montañola-Sales, X. Rubio-Campillo, J. Casanovas-Garcia, J. M. Cela-Espínand A. Kaplan-Marcusán (2014). *Interdisciplinary Applications of Agent-Based Social Simulation and Modeling (pp. 106-123)*. www.irma-international.org/chapter/large-scale-social-simulation-dealing-with-complexity-challenges-in-high-performance-environments/106764

A Multiphysics Based Finite Element Approach to Evaluate the Reliability of IC Packages

Padmanabhan Krishnan (2016). *Handbook of Research on Advanced Computational Techniques for Simulation-Based Engineering (pp. 71-102).*

www.irma-international.org/chapter/a-multiphysics-based-finite-element-approach-to-evaluate-the-reliability-of-ic-packages/140385

Information Processing by Chemical Reaction-Diffusion Media: From Computing to Vision

Nicholas G. Rambidi (2011). Computational Modeling and Simulation of Intellect: Current State and Future Perspectives (pp. 455-483).

www.irma-international.org/chapter/information-processing-chemical-reaction-diffusion/53316

Augmented Reality for Collaborative Assembly Design in Manufacturing Sector

Rui Chen, Xiangyu Wangand Lei Hou (2011). Virtual Technologies for Business and Industrial Applications: Innovative and Synergistic Approaches (pp. 105-117).

www.irma-international.org/chapter/augmented-reality-collaborative-assembly-design/43406

MLVQ: A Modified Learning Vector Quantization Algorithm for Identifying Centroids of Fuzzy Membership Functions

Kai Keng Angand Chai Quek (2011). Computational Modeling and Simulation of Intellect: Current State and Future Perspectives (pp. 485-509).

www.irma-international.org/chapter/mlvq-modified-learning-vector-quantization/53317