

## Chapter 2.8

# Design of an Enhanced 3G–Based Mobile Healthcare System

**José Ruiz Mas**

*University of Zaragoza, Spain*

**Eduardo Antonio Viruete Navarro**

*University of Zaragoza, Spain*

**Carolina Hernández Ramos**

*University of Zaragoza, Spain*

**Álvaro Alesanco Iglesias**

*University of Zaragoza, Spain*

**Julián Fernández Navajas**

*University of Zaragoza, Spain*

**Antonio Valdovinos Bardají**

*University of Zaragoza, Spain*

**Robert S. H. Istepanian**

*Kingston University, UK*

**José García Moros**

*University of Zaragoza, Spain*

### ABSTRACT

An enhanced mobile healthcare multi-collaborative system operating over Third Generation (3G) mobile networks is presented. This chapter describes the design and use of this system in different medical and critical emergency scenarios provided with universal mobile telecommunications system (UMTS) accesses. In these environments, it is designed to communicate healthcare personnel with medical specialists in a remote hospital. The system architecture is based on advanced signalling protocols that allow multimedia multi-collaborative conferences in IPv4/IPv6 3G scenarios. The system offers real-time transmission of medical data and vid-

eoconferencing, together with other non real-time services. It has been optimized specifically to operate over 3G mobile networks using the most appropriate codecs. Evaluation results show a reliable performance over IPv4 UMTS accesses (64 Kbps in the uplink). In the future, advances in m-Health systems will make easier for mobile patients to interactively get the medical attention and advice they need.

### INTRODUCTION

Mobile health (m-health) is an emerging area of telemedicine in which the recent development in mobile networks and telemedicine applications

converge. m-health involves the exploitation of mobile telecommunication and multimedia technologies and their integration into new mobile healthcare delivery systems (Istepanian & Lacal, 2003). Wireless and mobile networks have brought about new possibilities in the field of telemedicine thanks to the wide coverage provided by cellular networks and the possibility of serving moving vehicles. One of the first wireless telemedical systems that utilized second-generation (2G) global system for mobile communications (GSM) networks addressed the electrocardiogram (ECG) transmission issues (Istepanian, 2001a). In recent years, several m-health and wireless telemedical systems based on GSM were reported (Istepanian, 2001b), allowing the accomplishment of remote diagnosis in mobile environments, as well as communication to geographic zones inaccessible by wired networks. The recent developments in digital mobile telephonic technologies (and their impact on mobility issues in different telemedical and telecare applications) are clearly reflected in the fast growing commercial domain of mobile telemedical services. A comprehensive review of wireless telemedicine applications and more recent advances on m-health systems is presented in Istepanian, Laxminarayan, and Pattichis (2005).

However, 2G-based systems lack the necessary bandwidth to transmit bandwidth-demanding medical data. The third-generation (3G) universal mobile telecommunications system (UMTS) overcomes limitations of first and second mobile network generations supporting a large variety of services with different quality of service (QoS) requirements. However, this fact makes network design and management much more complex. New applications require networks to be able to handle services with variable traffic conditions keeping the efficiency in the network resources utilization. The UMTS air interface is able to cope with variable and asymmetric bit rates, up to 2 Mbps and 384 kbps in indoor and outdoor environments, respectively, with different QoS requirements

such as multimedia services with bandwidth on demand (Laiho, Wacker, & Novosad, 2000). In this kind of scenario, the emergence of 3G mobile wireless networks will permit to extend the use of m-health applications thanks to the provided higher transmission rates and flexibility over previous mobile technologies.

UMTS introduces the IP multimedia core network subsystem (IMS) (3GPP, 2005a), an IPv6 network domain designed to provide appropriate support for real-time multimedia services, independence from the access technologies and flexibility via a separation of access, transport and control. The fundamental reason for using IPv6 is the exhaustion of IPv4 addresses. Support for IPv4 is optional, but since network components require backward compatibility, it is clear that a dual stack configuration (IPv4 and IPv6) must be provided. The IMS uses the session initiation protocol (SIP) as signalling and session control protocol (Rosenberg et al., 2002). SIP allows operators to integrate real-time multimedia services over multiple access technologies such as general packet radio service (GPRS), UMTS or, ultimately, other wireless or even fixed network technologies (interworking multimedia domains). This chapter presents a 3G-based m-health system designed for different critical and emergency medical scenarios, as shown in Figure 1. Several medical specialists in the hospital take part in a multipoint conference with the ambulance personnel, receiving compressed and coded biomedical information from the patient, making it possible for them to assist in the diagnosis prior to its reception.

The 3G system software architecture includes intelligent modules such as information compression and coding, and QoS control to significantly improve transmission efficiency, thus optimizing the use of the scarce and variable wireless channel bandwidth compared to previous systems (Chu & Ganz, 2004; Curry & Harrop, 1998). Finally, unlike Chu and Ganz (2004), this m-health system follows a multi-collaborative design which supports IPv6/IPv4 interworking, uses SIP as

10 more pages are available in the full version of this document, which may be purchased using the "Add to Cart" button on the publisher's webpage: [www.igi-global.com/chapter/design-enhanced-based-mobile-healthcare/26235](http://www.igi-global.com/chapter/design-enhanced-based-mobile-healthcare/26235)

## Related Content

---

### Infant Cry Detection and Pain Scale Assessment: A Pilot Study

N. Sriraam and S. Tejaswini (2014). *International Journal of Biomedical and Clinical Engineering* (pp. 42-51). [www.irma-international.org/article/infant-cry-detection-and-pain-scale-assessment/115884](http://www.irma-international.org/article/infant-cry-detection-and-pain-scale-assessment/115884)

### Case Based Reasoning for Customizing Treatment Processes

Carolyn Kaiser (2009). *Handbook of Research on Distributed Medical Informatics and E-Health* (pp. 351-366). [www.irma-international.org/chapter/case-based-reasoning-customizing-treatment/19945](http://www.irma-international.org/chapter/case-based-reasoning-customizing-treatment/19945)

### Catering for Personal Autonomy with E-Mentoring Supported by Recommendations

Olga Santos, Carmen Barrera, Emanuela Mazzone and Jesus Boticario (2011). *Handbook of Research on Personal Autonomy Technologies and Disability Informatics* (pp. 191-203). [www.irma-international.org/chapter/catering-personal-autonomy-mentoring-supported/48282](http://www.irma-international.org/chapter/catering-personal-autonomy-mentoring-supported/48282)

### A Primitive Survey on Ultrasonic Imaging-Oriented Segmentation Techniques for Detection of Fetal Cardiac Chambers

Punya Prabha V. and Sriraam N. (2019). *International Journal of Biomedical and Clinical Engineering* (pp. 69-79). [www.irma-international.org/article/a-primitive-survey-on-ultrasonic-imaging-oriented-segmentation-techniques-for-detection-of-fetal-cardiac-chambers/233543](http://www.irma-international.org/article/a-primitive-survey-on-ultrasonic-imaging-oriented-segmentation-techniques-for-detection-of-fetal-cardiac-chambers/233543)

### Non-Contact Pulse Monitoring Using Live Imaging

Yuji Ohta and Miki Uchida (2013). *Technological Advancements in Biomedicine for Healthcare Applications* (pp. 240-246). [www.irma-international.org/chapter/non-contact-pulse-monitoring-using/70867](http://www.irma-international.org/chapter/non-contact-pulse-monitoring-using/70867)